

Annual Laboratory Control Sample

Date:1/22/2018

Samplers: Pete Hawes
BJ Barbrow

Calibration Buffers used

pH	Brand	Lot	Exp
pH 4	Oakton	VV2C	5/2019
pH 7	Oakton	VY1D	2/2019
pH 10	Oakton	VV1D	5/2019

Lab Control Sample Buffer used

pH	Brand	Lot	Exp
pH 7	Hydrion ¹	208010B	3/2019

Results

	pH	Temp
Pete Hawes	6.96	22.4
Bj Barbrow	6.96	22.6

¹ Capsulated powder is emptied into 100ML of Distilled water