



Enviro Compliance Laboratories, Inc.
 Post Office Box 919
 Verona, Virginia 24482
 (540) 248-1311
 Fax: (540) 248-4604
 www.envirocompliance.com
 eclverona@envirocompliance.com

Analytical Results

V6B12247

Allied Concrete Company
 Attn: Bill Shipe
 1000 Harris Street
 Charlottesville, VA 22903

Project No. : 2016
 Project Name : Annual
 Date Received: November 29, 2016
 Date Issued : December 05, 2016

Lab # V6B12247 - 1(A)/Sample ID : pH meter

Parameter	Result	Units	QL	Date/Time Prepared	Date/Time Analyzed	Method	Analyst
Thermometer Calibration							
NIST SN#1068	20.0		--	08-12/1000	08-12/1000		TGH
Client Therm	21.2		--	08-12/1000	08-12/1000		TGH
Correction	-1.2		--	08-12/1000	08-12/1000		TGH
NIST SN#1068	4.0		--	08-12/1000	08-12/1000		
Client Therm	3.6		--	08-12/1000	08-12/1000		
Correction	+0.4		--	08-12/1000	08-12/1000		

V6B12247

V6B12247

Tamara G. Hostetter

Tamara G. Hostetter
 Laboratory Manager

Report #: V6B12247 Page 1 of 2

© EnviroCompliance 2016
 This report is only valid as the complete digital document authored by EnviroCompliance Labs, Inc.





Report Annex

Enviro Compliance Laboratories, Inc.
Post Office Box 919
Verona, Virginia 24482
(540) 248-1311
Fax: (540) 248-4604
www.envirocompliance.com
eclverona@envirocompliance.com

Abbreviations:

NR = Not Reported

ND = Not Detected

BQL = Below Quantitation Level (Result is less than stated QL or <QL)

< = Result is less than Quantitation Limit

J = Result is estimated outside of calibration range

Quality Assurance Flags:

L = LCS did not meet method criteria.

HT = Sample was not analyzed/received within holding time.

T = Sample was not received at appropriate temperature (<6.0C).

P = Sample was not properly preserved or received in inappropriate container.

R = Corr Coef <.995

C = Initial Instrument Calibration (Second Source) did not meet criteria

V = Continuing Calibration Verification did not meet criteria

S = Matrix Spike did not meet criteria

D = Duplicate did not meet criteria

H/B = Blank did not meet QC criteria

SR = Surrogate Recovery was not in acceptable limits.

TOX = Toxicity exhibited in BOD

G = GGA/Int. QC was not 198.5+/-30.5

Y = Yield not within 2-200mg

Cl = Residual chlorine was detected in the micro sample >15mg/l.

Micro methods do not perform properly for samples with residual chlorine.

* = Analysis was subcontracted

** = Non-accreditable/non-accredited parameter

Notes:

Analysis was performed in accordance to NELAC requirements unless otherwise noted.

All methods are 40 CFR 136 May 18, 2012, Table IB approved.

Report #: V6B12247 Page 2 of 2

© EnviroCompliance 2016

This report is only valid as the complete digital document authored by EnviroCompliance Labs, Inc.

