



Certificate of Calibration

DATE CALIBRATED: 8/9/17

SITE NAME: ALLIED Concrete

METER TYPE AND MODEL: PH Meter - PH TESTR 30

METER SERIAL NUMBER: 863810

PROBE SERIAL NUMBER: N/A

THERMOMETER NUMBER: 2136892

NIST TEMP READING:	<u>8.4°C</u>	<u>21.8°C</u>	<u>29.7°C</u>
METER TEMP READING:	<u>8.1°C</u>	<u>20.8°C</u>	<u>28.4°C</u>
CORRECTION FACTOR:	<u>+0.3°C</u>	<u>+1.00°C</u>	<u>+1.3°C</u>

SIGNATURE OF CALIBRATOR: [Signature]

NOTE: Meter PASSED Temperature Calibration.
Due to high CORRECTION FACTOR REPLACING
METER IS RECOMMENDED.